

# INSTRUCTION MANUAL

## **EDX-15A**

### Compact Recording System

## LOW POWER STRAIN/VOLTAGE MEASURING UNIT

Thank you for purchasing KYOWA's product EDX-15A Low Power Strain/Voltage Measuring Unit.

Read this Instruction Manual carefully in order to make full use of the high performance capabilities of the product.

Do not use the product in methods other than described in this Manual.

This Manual only describes hardware operation of the EDX-15A. For the Dynamic Data Recording Software DCS-100A, see the DCS-100A Instruction Manual (For EDX-10A/B Operation).

Generally, company names and trade names described in this Instruction Manual are trademarks or registered trademarks of the companies.

This Instruction Manual may not be copied or reproduced, in whole or part, without consent of KYOWA.

Copyright © Kyowa Electronic Instruments Co., Ltd. All rights reserved.

The contents of the Instruction Manual are subjected to change without prior notice.

We assume no responsibility whatever for any damage resulting from the use of this product.



# CONTENTS

|                                                                                                         |    |
|---------------------------------------------------------------------------------------------------------|----|
| CONTENTS .....                                                                                          | 1  |
| STANDARD ACCESSORIES .....                                                                              | 2  |
| OPTIONAL ACCESSORIES .....                                                                              | 2  |
| SAFETY PRECAUTIONS (Do not forget to read the safety precautions prior to use.) .....                   | 3  |
| LIMITED LIFE PARTS AND PREVENTIVE MAINTENANCE .....                                                     | 6  |
| EDX-15A REPLACEMENT BEFORE THE END OF SERVICE LIFE .....                                                | 6  |
| NOTATIONS USED IN THE INSTRUCTION MANUAL .....                                                          | 6  |
| IMPORTANT PRECAUTIONS WHEN USING EDX-15A .....                                                          | 7  |
| CE MARKING .....                                                                                        | 7  |
| SURFACE APPEARANCE .....                                                                                | 7  |
| 1. OUTLINE OF THE PRODUCT .....                                                                         | 8  |
| 1-1. OUTLINE OF THE PRODUCT .....                                                                       | 8  |
| 1-2. FEATURES .....                                                                                     | 8  |
| 1-3. MODELS .....                                                                                       | 8  |
| 2. CONTROLS AND INDICATORS .....                                                                        | 9  |
| 2-1. FRONT .....                                                                                        | 9  |
| 2-2. TOP .....                                                                                          | 9  |
| 2-3. SIDE .....                                                                                         | 10 |
| 2-4. BOTTOM .....                                                                                       | 10 |
| 3. PIN ASSIGNMENT AND CONNECTION OF INPUT CONNECTOR .....                                               | 11 |
| 3-1. PIN ASSIGNMENT OF INPUT CONNECTORS .....                                                           | 11 |
| 3-2. CONNECT THE MEASURING UNITS .....                                                                  | 12 |
| 4. CONNECT A STRAIN GAGE TRANSDUCER .....                                                               | 13 |
| 4-1. When using the accessory input cable (U-124) .....                                                 | 13 |
| 4-2. When using the optional accessory input adapter (UI-51A) .....                                     | 14 |
| 5. CONNECT VOLTAGE .....                                                                                | 15 |
| 5-1. When using the standard accessory input cable (U-124) and voltage conversion adapter (FV-1A) ..... | 15 |
| 5-2. When using the optional accessory input cable (U-125) .....                                        | 17 |
| 5-3. When using the accessory input adapter (UI-51A) .....                                              | 18 |
| 6. TROUBLE SHOOTING .....                                                                               | 20 |
| 7. MAINTENANCE .....                                                                                    | 21 |
| 8. SPECIFICATIONS .....                                                                                 | 22 |
| 9. OUTSIDE DRAWING .....                                                                                | 23 |

## STANDARD ACCESSORIES

The following accessories are packed with the EDX-15A.

When unpacking, check the contents to ensure that all accessories are enclosed.

|                                                        |         |
|--------------------------------------------------------|---------|
| Input cable U-124: .....                               | 1       |
| Voltage conversion adapter FV-1A: .....                | 4       |
| Rubber Foot: .....                                     | 8/sheet |
| Screws (cross recessed head): .....                    | 2       |
| Term and Conditions of .....                           | 1       |
| Product Warranty:                                      |         |
| EDX-15A READ IN THE BEGGNING (English/ Japanese) ..... | 1       |
| EDX-15A READ IN THE BEGGNING (Chinese) .....           | 1       |

## OPTIONAL ACCESSORIES

|                                                      |                |
|------------------------------------------------------|----------------|
| Input adapter: .....                                 | UI-51A         |
| One-touch type input adapter: .....                  | UI-52A         |
| The bridge adapter for Quarter bridge system : ..... | UI-53A-120/350 |
| The bridge adapter for Quarter bridge system: .....  | UI-54A-120/350 |
| Quick-fitting input adapter: .....                   | UI-55A         |
| BNC input cable: .....                               | U-125          |
| Bridge box coupling cable: .....                     | U-126 (50cm)   |
| Input Connector Set: .....                           | EDX10-DSUB     |

## SAFETY PRECAUTIONS (Do not forget to read the safety precautions prior to use.)

The EDX-15A is designed to be used according to "8. SPECIFICATIONS."

Do not use the EDX-15A in an environment exceeding the specifications.

Or, system failure may result in.

### PRIOR TO USE

For safe use of the EDX-15A, do not forget to read the "Safety Precautions" prior to use.

Kyowa Electronic Instruments Co., Ltd. assumes no liability for any damage resulting from the user's failure to comply with the safety precautions.

For safety operation of the EDX-15A, the following symbol mark is attached to the EDX-15A.

|                                                                                   |                                                                                               |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------|
|  | Indicates "PROTECTIVE GROUND TERMINAL."<br>Be sure to connect the GND terminal to the ground. |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------|

### SAFETY SYMBOLS

For safety operation of the EDX-15A, the following symbol marks are used in the Instruction Manual.

|                                                                                                   |                                                                                                          |
|---------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------------------------|
|  <b>WARNING</b>  | Improper operation of the system may result in death or severe injury of the operator.                   |
|  <b>CAUTION</b> | Improper operation of the system may result in injury of the operator and physical damage of the system. |

|                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                    |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <div> <b>WARNING</b></div> <div><ul style="list-style-type: none"><li>• <b>WARNING</b><br/>Be sure to observe the warning and precautions described in the "EDX-10B Instruction Manual (FOR HARDWARE)" and "EDX-15A Instruction Manual".</li><li>• <b>Avoid using in environment with inflammable gas, etc.</b><br/>To prevent the risk of fire hazard or explosion, do not use the EDX-15A in environment with inflammable gas, vapor, or dust.</li><li>• <b>If any trouble occurs</b><br/>To prevent fire hazards, if smoke is emerging from the EDX-15A, immediately disconnect the USB cable or AC adapter to terminate the system operation.</li><li>• <b>Compatible input</b><br/>The EDX-15A is a measuring instrument for a strain gage transducer and voltage.<br/>When measuring strain gage transducer, be sure to use the strain gage transducers and bridge boxes only.<br/>Failure to do so will pose the risk of electric shock, deteriorated performance and system failure.<br/>When measuring voltage, be sure to input voltage only.<br/>Failure to do so will pose the risk of electric shock, deteriorated performance and system failure.</li></ul></div> |
|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

## CAUTION

### • CAUTION

Be sure to observe the caution and precautions described in the “EDX-10B Instruction Manual (FOR HARDWARE)” and “EDX-15A Instruction Manual”.

### • Do not use the EDX-15A outdoors.

Or, it may cause electric shock, fire hazard, lower the performance and cause trouble.

### • Use the EDX-15A within temperature ranging from 0 to 40°C.

Use at temperatures exceeding the specified range may lower the performance and cause trouble.

If use under direct sunlight or in a cold place is inevitable, prepare a sunscreen or take proper measures to keep it warm.

### • Use the EDX-15A in the specified operating humidity of 20 to 85%.

Use in a humid place exceeding the specified range or where it is exposed to splashing water may lower the performance and cause trouble.

### • Do not use the EDX-15A immediately after the change in the environment.

Leave the EDX-15A as it is to acclimate to the environment.

Abrupt change in ambient temperature due to transportation, etc. may cause dew condensation, which may result in lower performance and troubles.

### • Do not use the EDX-14A where it is subject to significant vibration or impact.

The vibration resistance and impact resistance are as follows.

Vibration resistance: 29.42m/s<sup>2</sup> (3G), 5 to 200Hz (when operating)

Impact Resistance: 294m/s<sup>2</sup>(30G)/11msec

Use the EDX-15A in an area where vibration and impact can be kept within the scope of specifications.

Continuous vibration or severe impact may cause deteriorated performance and system failure.

### • Do not use the EDX-15A in strong electromagnetic field.

Use the EDX-15A in an environment where magnetic field is acceptable to the PC.

Performance may be lowered and erroneous operation and troubles may result if it is used near a telemetry system, microwave oven, electronic furnace or any other equipment generating a strong magnetic field.

### • Do not pull cables.

Lay cords and cables with a certain allowance so that unreasonable force is applied to the connections.

Pulling or unreasonable force may cause troubles or interrupt the measurement.

### • Avoid installing sensors and EDX-15A near a welding machine.

Failure to do so will pose the risk of erroneous data, malfunction and failure.

### • Do not disassemble or remodel the EDX-15A.

Disassembling or remodeling by user may cause electric shock hazards or damage the EDX-15A.

This warranty does not cover any damaged or defective parts that results from disassembling or remodeling.

### • Do not use the EDX-15A under dusty environment.

This can lead to performance problems and decreased operating efficiency.

Be careful that dust may not enter the EDX-15A not only during operation but also under stored conditions.

### • Preheat the EDX-15A before use.

After the power ON, always preheat the EDX-15A for approximately 10 minutes.

### •Cleaning.

When the product gets dirty, clean the product with a dry soft cloth. When dust exists inside the EDX-15A, clean it by using an air blow gun. Do not touch electronic parts.

### •Be sure to use the accessory screws (binding head screw, M3 × 5) for coupling measuring units.

Longer screws will damage the inner PCB components.

## CAUTION

- Do not apply any voltage beyond the specified input voltage range.  
Failure to do so will pose the risk of deteriorated performance and system failure.  
See the following precautions before measuring.
  - 1) Use a tester or the like to measure the voltage between negative sides of the DC voltage and confirm that the voltage is approximately 0 V.
  - 2) Confirm that the voltage between the negative side of the DC voltage and the GND terminal of the EDX-10A is approximately 0 V.
  - 3) When connecting an input, correctly connect the positive and negative sides.  
Or, internal circuits have short-circuited and cause troubles.
  - 4) You cannot measure the middle voltage of the battery connected with the sensors.  
Or, internal circuits have short-circuited and cause troubles.Failure to comply with the above precautions will result in a short circuit of the signal source (e.g. DC power supply) of the measuring target, thus causing system failure and a fire hazard.

## LIMITED LIFE PARTS AND PREVENTIVE MAINTENANCE

The EDX-15A consists of various electronic components and those components, if not all, have a useful life.

Using them exceeding the years specified according to each part type (useful life) may affect the characteristics of the EDX-15A, resulting in a malfunction or a failure.

You need to replace parts with a regular preventive maintenance schedule.

Limited life parts used with the EDX-15A are as follows.

○ Aluminum electrolytic capacitor

The signal-noise ratio will lower due to capacity low or smoke will be emitted due to liquid leak, resulting in a malfunction of the EDX-15A.

○ AC adapter (optional accessories)

Heat and smoke will be emitted, resulting in an unstable voltage.

To operate the EDX-15A normally, finding a sign of the EDX-15A failure early by daily/periodic inspections and taking the corrective action are required.

For inspection, contact KYOWA or our representatives.

\* All components do age and will fail.

## EDX-15A REPLACEMENT BEFORE THE END OF SERVICE LIFE

Preventive maintenance and replacement parts are a cost-effective way of keeping the performance and extending the service life of the EDX-15A.

Regardless of the replacement parts, the EDX-15A itself gradually deteriorates with age.

Before the expected service life is reached, consider replacing the EDX-15A with a new one or the latest series as preventive maintenance.

## NOTATIONS USED IN THE INSTRUCTION MANUAL

### Informational Notes

Certain notations are used as necessary to attract your attention to information that requires special care when handling the EDX-15A, and to information provided for reference purposes.

### Examples of Notations

**NOTE**

Essential precautions required when handling the EDX-15A.

**MEMO**

Reference items required when handling the EDX-15A.

## IMPORTANT PRECAUTIONS WHEN USING EDX-15A

### NOTE

1. Be sure to use the accessory screws (binding head screw,  $M3 \times 5$ ) for coupling measuring units.  
Longer screws will damage the inner PCB components.
2. After the power ON, always preheat the EDX-15A for approximately 10 minutes.

## CE MARKING

### NOTE

1. Be sure to use the input cable (U-124) for the EDX-15A. Wrap the ferrite core two turns around every channel cable.  
Recommended ferrite core                      Model: E04SR170730A  
Seiwa Electric Mfg. Co.,Ltd.
2. Use a shielded cable between the input cable (U-124) and strain gage transducer.

## SURFACE APPEARANCE

- Some products have a weld line often looks like a scratch or a line on the product appearance. It is called a “weld line” which occurred during parts molding. Weld lines are not a crack or scratch. There are no problems with the quality, performance and strength of the product.

# 1. OUTLINE OF THE PRODUCT

## 1-1. OUTLINE OF THE PRODUCT

The EDX-15A is a measuring unit for a strain gage transducer and voltage. The EDX-15A has to be installed in the EDX-10B.

One EDX-15A is capable of measurement of 4 channels and maximum 16 channels with four EDX-15A units.

When connecting to EDX-10B, maximum 16 channels with four measuring unit can operate on USB bus power. (When used USB3.0 port. If you use USB2.0, then maximum 8 channels with two EDX-15A units. )

EDX-15A offers additional convenience for measurements because they are powered from the USB interface and do not require a separate power supply.

## 1-2. FEATURES

- Compact and lightweight
- Stacking connectors enable easy connection and make synchronous cables unnecessary.
- Since one measuring unit is capable of measurement of 4 channels and maximum 16 channels with four measuring units, the EDX-10B is compatible for measurement in a wide range from a simple experiment to advanced measurement.
- When used USB3.0 port, maximum 16 channels with 4 measuring unit can operate on USB bus power.
- A/D conversion function (24-bit)
- Strain measurement can measure maximum 50,000  $\mu\text{m/m}$  at strain measurement, voltage measurement can measure maximum 50V.
- Capable of checking the ZERO balance of the bridge box and strain gage transducer by selecting the "Balance" [NONE] on the channel condition setting window.

## 1-3. MODELS

EDX-10series composed control unit EDX-10B and measuring units. There are five types of measuring unit.

EDX-11A that measure with strain gage type transducers, load, pressure, displacement, etc. EDX-14A is low power strain measuring unit. EDX-12A is voltage measuring unit, which can measuring Voltage. EDX-15A is low power strain / voltage measuring unit, which can measuring strain and voltage.


EDX-13A is thermocouple measuring unit, which can measuring temperature by thermocouple.

Table 1 EDX-10series models

| Models  | Name                                      |
|---------|-------------------------------------------|
| EDX-10B | Control Unit                              |
| EDX-11A | Strain measuring unit                     |
| EDX-12A | Voltage measuring unit                    |
| EDX-13A | Thermocouple measuring unit               |
| EDX-14A | Low power strain measuring unit           |
| EDX-15A | Low power strain / voltage measuring unit |


## 2. CONTROLS AND INDICATORS

### 2-1. FRONT




- (1) Input Connector      Input connector for 4 channels.  
Connect the accessory input cable (U-124) or optional accessory input adapter (UI-51A).etc.
- (2) Fixing metal      For fixing the D-sub connector of the input cable (U-124).

### 2-2. TOP



- (3) Unit-coupling Connector      Connector for connecting the EDX-10B or measuring units on the EDX-15A.


## 2-3. SIDE




- (4) Clearance hole For fixing the EDX-10B or measuring unit on the EDX-15A by using a screw.
- (5) Finger slot For connecting/disconnecting the EDX-10B and a measuring unit.
- (6) Rubber foot slot The EDX-15A has small rubber feet (optional) to keep it from sliding around when placed on a flat surface.
- (7) Fixing screw hole For fixing the measuring unit, under the EDX-15A, by using a screw.

## 2-4. BOTTOM

With Cap



With no Cap




- (8) Rubber foot slot The EDX-15A has small rubber feet (optional) to keep it from sliding around when placed on a flat surface.
- (9) Cap Cap for protecting the unit-coupling connector on the EDX-15A bottom. Remove the cap before connecting measuring units.
- (10) Unit-coupling Connector Connector for connecting measuring units under the EDX-15A.

### 3. PIN ASSIGNMENT AND CONNECTION OF INPUT CONNECTOR

#### 3-1. PIN ASSIGNMENT OF INPUT CONNECTORS

| Pin No. | Signal Name | Function | Pin No. | Signal Name | Function |
|---------|-------------|----------|---------|-------------|----------|
| 1       | CH1 A       | CH1 +BV  | 20      | CH1 B       | CH1 -IN  |
| 2       | NC          |          | 21      | CH1 D       | CH1 +IN  |
| 3       | CH1 C       | CH1 -BV  | 22      | NC          |          |
| 4       | NC          |          | 23      | NC          |          |
| 5       | NC          |          | 24      | CH2 B       | CH2 -IN  |
| 6       | CH2 A       | CH2 +BV  | 25      | CH2 D       | CH2 +IN  |
| 7       | NC          |          | 26      | NC          |          |
| 8       | CH2 C       | CH2 -BV  | 27      | NC          |          |
| 9       | NC          |          | 28      | NC          |          |
| 10      | CH3 A       | CH3 +BV  | 29      | CH3 B       | CH3 -IN  |
| 11      | NC          |          | 30      | CH3 D       | CH3 +IN  |
| 12      | CH3 C       | CH3 -BV  | 31      | NC          |          |
| 13      | NC          |          | 32      | NC          |          |
| 14      | NC          |          | 33      | CH4 B       | CH4 -IN  |
| 15      | CH4 A       | CH4 +BV  | 34      | CH4 D       | CH4 +IN  |
| 16      | NC          |          | 35      | NC          |          |
| 17      | CH4 C       | CH4 -BV  | 36      | NC          |          |
| 18      | NC          |          | 37      | NC          |          |
| 19      | NC          |          |         |             |          |



#### NOTE

- [NC]: Non-connection pin  
Do not connect anything. Or, deterioration of the instrument may result.
- Connect the shield to the D-sub connector case.


### 3-2. CONNECT THE MEASURING UNITS

This section describes how to connect the EDX-10B and measuring units into the EDX-15A.


#### NOTE

- Mount the cap to the unit-coupling connector of the very-bottom measuring unit.  
**Or, it may cause system failures and accidents.**
- When used the optional AC adapter, you can connect the measuring units up to 4 units.  
Do not connect 5 or more measuring units.
- When used USB bus power, the number of connectable measuring unit is limited.  
For details, “EDX-10B Instruction Manual (FOR HARDWARE) 3-2-1.NUMBER OF CONNECTABLE MEASURING UNIT”. (If you use EDX-10A, the number of connectable measuring unit is one.)

- 1) To connect the EDX-10B or a measuring unit on the EDX-15A, remove the cap of the EDX-10B or a measuring unit to be connected.



- 2) Connect the unit-coupling connector on the EDX-10B bottom or on the measuring unit bottom and connector on the EDX-15A top. Use the finger slots to connect the EDX-10B or measuring unit and EDX-15A.




#### NOTE

- Do not bend or twist the EDX-10B and the measuring unit for connection.  
**Or, the holding pin and connectors on the measuring unit will be damaged.**

- 3) To connect a measuring unit under the EDX-15A, remove the cap of the EDX-15A.

- 4) Connect the unit-coupling connector on the EDX-15A bottom and connector on the measuring unit top.

- 5) Fasten the accessory screws on both sides of the measuring units.



#### NOTE

- Be sure to use the accessory screws (binding head screw, M3 × 5).  
Longer screws will damage the inner PCB components.
- Tightening torque: 0.7 Nm or less

## 4. CONNECT A STRAIN GAGE TRANSDUCER

The EDX-15A is a measuring instrument for a strain gage transducer. The input cable should be connected as follows.


### NOTE

Before connecting a sensor, disconnect the USB cable and AC adapter from the EDX-10B to turn OFF the power.


#### 4-1. When using the accessory input cable (U-124)

- 1) When measuring the output of the strain gage transducer, connect the output connector of the strain gage transducer to the D-sub connector of the input cable (U-124).

When measuring the output of the bridge box, connect the output connector of the bridge box to the D-sub connector of the input cable (U-124).



- 2) Insert the D-sub connector case into the input connector of the EDX-15A.
- 3) Fasten the D-sub connector case tightly with the two screws.
- 4) To disconnect, unfasten the two screws on the D-sub connector case. Hold the D-sub connector case and pull it.



### MEMO

Before recording data, select the strain mode and monitor the noise.


If excessive noise is measured, ground the shield of the output cable or connect it to the GND terminal of the EDX-10B, whichever results in less noise.

### NOTE

- When you disconnect the input cable, do not pull on the cable itself.  
Failure to do so will pose the risk of disconnection and connector failure.
- Transducer with a remote sensing function are not available.
- Do not fasten the screws on the D-sub connector case too tightly. Or, the screws will be damaged.

## 4-2. When using the optional accessory input adapter (UI-51A)

- 1) When the cable of the strain gage transducer has bare wires, solder the bare wires onto the input adapter tab as follows.



### NOTE

- Do not place the input adapter on any metal surface. Do not touch the tab directly with hands.  
Failure to do so will pose the risk of failure, electric shock, and erroneous data.
- When you disconnect the input adapter, do not pull on the cable itself.  
Failure to do so will pose the risk of disconnection and tab failure.

## 5. CONNECT VOLTAGE


The EDX-15A is a measuring instrument for voltage. The input cable should be connected as follows.

### NOTE


Before connecting voltage, disconnect the USB cable and AC adapter from the EDX-10B to turn OFF the power.

### 5-1. When using the standard accessory input cable (U-124) and voltage conversion adapter (FV-1A)


- 1) When measuring the voltage, connect the voltage conversion adapter (FV-1A) to the input cable (U-124) as lower figure.



- 2) Voltage output BNC connector connect to the voltage conversion adapter (FV-1A) as lower figure.



- 3) Insert the D-sub connector case into the input connector of the EDX-15A.
- 4) Fasten the D-sub connector case tightly with the two screws.
- 5) To disconnect, unfasten the two screws on the D-sub connector case. Hold the D-sub connector case and pull it.



#### MEMO

Before recording data, select the voltage mode and monitor the noise.


If excessive noise is measured, ground the shield of the output cable or connect it to the GND terminal of the EDX-10B, whichever results in less noise.

#### NOTE


- When you disconnect the input cable, do not pull on the cable itself.  
Failure to do so will pose the risk of disconnection and connector failure.
- Do not fasten the screws on the D-sub connector case too tightly. Or, the screws will be damaged.

## 5-2. When using the optional accessory input cable (U-125)

- 1) Connect the BNC connector to the D-sub connector of the input cable (U-125).



- 2) Insert the D-sub connector case into the input connector of the EDX-15A.
- 3) Fasten the D-sub connector case tightly with the two screws.
- 4) To disconnect, unfasten the two screws on the D-sub connector case. Hold the D-sub connector case and pull it.



### MEMO

Before recording data, select the voltage mode and monitor the noise.


If excessive noise is measured, ground the shield of the output cable or connect it to the GND terminal of the EDX-10B, whichever results in less noise.

### NOTE

- When you disconnect the input cable, do not pull on the cable itself.  
Failure to do so will pose the risk of disconnection and connector failure.
- Do not fasten the screws on the D-sub connector case too tightly. Or, the screws will be damaged.

### 5-3. When using the accessory input adapter (UI-51A)

- 1) When the cable, connected to the battery has bare wires, solder the bare wires onto the input adapter tab as follows.  
Maintain the temperature of the soldering iron tip between 260 to 350°C (within 5 seconds).



#### MEMO

Before recording data, select the strain mode and monitor the noise.

If excessive noise is measured, ground the shield of the output cable or connect it to the GND terminal of the EDX-10B, whichever results in less noise.

#### NOTE

- Every CH is not insulated.
- Correctly connect the wiring.
  - Positive (+) input: D terminal
  - Negative (-) input: B terminal
- Special care should be taken when handling soldering irons and solders.
- Tabs can endure several times of soldering with temperature of the soldering iron tip from 260 to 350°C (within 5 seconds). If you disconnect and solder leadwires for several times, patterns on the tab may be damaged. In that case, stop using and purchase a new input adapter (UI-51A).
- Do not place the input adapter on any metal surface. Do not touch the tab directly with hands.  
Failure to do so will pose the risk of failure, electric shock, and erroneous data.
- When you disconnect the input adapter, do not pull on the cable itself.  
Failure to do so will pose the risk of disconnection and tab failure.

# WARNING

- Do not apply any voltage beyond the specified input voltage range.

The input voltage range of the EDX-15A is  $\pm 50\text{V}$ .


Failure to do so will pose the risk of electric shock, deteriorated performance and system failure.

- Precautions when measuring the DC voltage of a battery, power supply, etc.

See the following precautions when measuring the DC voltage of a battery, power supply, etc.

- 1) Use a tester or the like to measure the voltage between negative sides of the DC voltage and confirm that the voltage is approximately 0 V.
- 2) Confirm that the voltage between the negative side of the DC voltage and the GND terminal of the EDX-10B is approximately 0 V.
- 3) When connecting an input, correctly connect the positive and negative sides.  
Or, internal circuits have short-circuited and cause troubles.
- 4) You cannot measure the middle voltage of the battery connected with the sensors.  
Or, internal circuits have short-circuited and cause troubles.

Failure to comply with the above precautions will result in a short circuit of the signal source (e.g. DC power supply) of the measuring target, thus causing system failure and a fire hazard.



## 6. TROUBLE SHOOTING

When the EDX-15A does not operate as expected or when it operates unstably, trace the cause before concluding a defect of EDX-15A.

Items to be checked and countermeasure against various phenomena are described in the following table.

If troubles are not solved despite countermeasures, contact KYOWA or our representative.

If the EDX-15A is damaged due to disassembling or remodeling by the user, KYOWA may refuse the repair.

| Trouble                                      | Checked Items and Countermeasures                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                |
|----------------------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| The balancing operation cannot be conducted. | <ul style="list-style-type: none"> <li>• The "Balance" is [OFF] or [NONE] on the channel condition setting window.<br/>Set the "Balance" [ON]. If OK still does not appear, a sensor is not connected, a sensor is failure, or sensor input is unstable.</li> <li>• The input exceeds the half of the input range of the measuring unit.</li> <li>• The input has excessive noise. For details, see the following "Noisy" part.</li> </ul>                                                                                                                                                                                                                                       |
| Unstable zero point                          | <ul style="list-style-type: none"> <li>• The zero point will drift when the EDX-15A is not sufficiently preheated.<br/>Always preheat the EDX-15A for approximately 10 minutes.</li> <li>• The "no load output" of the sensor is unstable.<br/>Repair the cable or replace the sensor with a new one.</li> </ul> <p>If troubles are not solved despite countermeasures, contact KYOWA or our representative.</p>                                                                                                                                                                                                                                                                 |
| Noisy                                        | <ul style="list-style-type: none"> <li>• Be sure to connect the GND terminal to the ground.</li> <li>• Do not locate the input cable near the inverter motor, machine tool, welding machine, or AC motor power supply. Separately locate the input cable from the power source.<br/>For details, see the following MEMO.</li> </ul> <hr/> <ul style="list-style-type: none"> <li>• Set the low-pass filter to reduce noise.<br/>Set the cutoff frequency of the low-pass filter twice as large as the input signal frequency.<br/>Example:<br/>Suppose the input is DC to 20 Hz. Set the cutoff frequency 100 Hz.<br/>If the input signal rapidly changes, set 2k Hz.</li> </ul> |

### MEMO

#### Inductive interference

Using the EDX-15A, near equipment generating strong magnetic field, generates induction noise. The equipment includes large motors, transformer, and ferro-resonant constant voltage transformer. Locating the EDX-15A, far away from measuring points, also generates induction noise. Take the following measures to protect against induction noise.

- Separately locate the EDX-15A from the power source.
- Use a shielded cable for sensor connection. Connect the shield to the E terminal of the input cable (U-124).

#### Potential interference

Potential interference occurs when the target to be measured has a DC/AC potential.

At this time, grounding the EDX-15A is not necessarily a good thing.

It may be appropriate to connect the target to be measured and EDX-10B.

If there is induction noise or if the target to be measured has potential, it is rare to be solved by a simple method.

Especially nowadays, the number of electrical devices that are used together in the measuring environment are increasing resulting to affect with each other as the noise cause. The above mentioned countermeasures are mere examples. Fully consider the operational environment and take appropriate measures to obtain the most preferable results.

## 7. MAINTENANCE

For the EDX-15A, the following items should be serviced for scheduled maintenance.  
Maintenance service should be performed at the intervals specified below.

Table 6-1 Calibration of apparatus

| Item                                      | Content                                                                                                                                                                                                    | Service intervals |
|-------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------|
| Evaluation of accuracy of measuring units | An annual maintenance inspection (including a calibration and adjustment) is recommended to ensure the accuracy of measuring units.<br>Must be returned to KYOWA for the scheduled maintenance inspection. | Once in a year    |


Particularly if the EDX-15A is used continuously for long periods of time, early replacement of parts is required for the purposes of safety and stable operation.

## 8. SPECIFICATIONS

| No.  | Item                                     | Strain measurement                                                                                | Voltage measurement           |
|------|------------------------------------------|---------------------------------------------------------------------------------------------------|-------------------------------|
| (1)  | Name                                     | Low power strain/voltage measuring unit                                                           |                               |
| (2)  | Model                                    | EDX-15A                                                                                           |                               |
| (3)  | Compatible sensor                        | Strain gage transducer and strain gage (*1)                                                       | Voltage<br>(Unbalanced input) |
| (4)  | The number of input channels             | 4                                                                                                 |                               |
| (5)  | Measurement range                        | 10000, 50000 $\mu\text{m/m}$ (2 range)                                                            | 10, 50V (2 range)             |
| (6)  | Compatible bridge resistance             | 120 to 1 k $\Omega$                                                                               | ---                           |
| (7)  | Excitation voltage                       | 1 VDC                                                                                             | ---                           |
| (8)  | Gage factor                              | 2.00 fixed                                                                                        | ---                           |
| (9)  | Range accuracy                           | Within $\pm 0.3$ %FS                                                                              |                               |
| (10) | Balance adjustment range                 | Within $\pm 1/2$ FS of setting range                                                              |                               |
| (11) | Nonlinearity                             | Within $\pm 0.1$ %FS                                                                              |                               |
| (12) | AD converter resolution                  | 24 bit                                                                                            |                               |
| (13) | Frequency response range                 | DC to 2 kHz                                                                                       |                               |
| (14) | Low-pass filter                          | Cutoff frequency: 100, 2kHz<br>2nd order Butterworth                                              |                               |
| (15) | Operating temperature and Humidity range | 0 to 40 °C, 20 to 90% (non-condensing)                                                            |                               |
| (16) | Input connector                          | 37-pin D-sub connector                                                                            |                               |
| (17) | Power supply                             | 5 VDC supplied by control unit                                                                    |                               |
| (18) | Current consumption                      | 150mA or less<br>(120 $\Omega$ load with all channels connected, at power source 5VDC)            |                               |
| (19) | Weight                                   | Approx. 150g                                                                                      |                               |
| (20) | Dimensions                               | 84.0(W)×26.6(H)×84.0(D) mm (excluding protrusions)                                                |                               |
| (21) | Compliance                               | Directive 2014/30/EU (EMC)<br>Directive 2011/65/EU, (EU)2015/863(10 restricted substances) (RoHS) |                               |

(\*1) When using the strain gage will require additional the bridge box or the input adaptor.

9. OUTSIDE DRAWING



Unit: mm



**Kyowa Electronic Instruments Co.,Ltd.**

3-5-1, Chofugaoka, Chofu, Tokyo 182-8520 Japan

<https://www.kyowa-ei.com/>